**Table S1 - Data extracted from the studies evaluating the performance of CT antigen-based POCTs**

| Study | First Author | Publication Year | Location | Test evaluated | Specimen | Reference assay | confirm assay for discrepant results | Sample size | Population Prevalence (%) | Reported sensitivity (n/n) | Reported specificity (n/n) | PPV | NPV | DOR |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 1 | Yin YP | 2006 | China | Clearview Chlamydia MF | Vaginal swabs | Roche Amplicor CT assay | NA | 1497 females | 13.2 | 32.8% (66/201) | 99.2% (1285/1296) | 85.7% | 90.5% | 57 |
| | | | | Clearview Chlamydia MF | Cervical swabs | Roche Amplicor CT assay | NA | 1497 females | 13.2 | 49.7% (98/197) | 97.9% (1273/1300) | 78.4% | 92.8% | 47 |
| 2 | Mahilum-Tapay L | 2007 | UK | Chlamydia Rapid Test (CRT) | Vaginal swabs | Roche Amplicor CT assay and transcription medicated assay | Aptima Combo 2 transcription mediated amplication assay | 1349 females | 8.2 | 83.5% (91/109) | 98.9% (1224/1238) | 86.7% | 98.6% | 442 |
| 3 | Sabido M | 2009 | Guatemala | Chlamydia test card (Ultimed Products, GmbH, Germany) | Cervical swabs | Roche Amplicor CT/NG test | NA | 276 females | 9.8 | 62.96% (17/27) | 99.6% (248/249) | 94.4% | 96.1% | 422 |
| 4 | Nadala EC | 2009 | UK | Chlamydia Rapid Test (CRT) | First void male urine samples | Roche Diagnostic Systems, Branchburg, NJ; Gen-Probe, San Diego, CA | Aptima CT kit | 1211 males | 9.0 | 82.6% (90/109) | 98.5% (1085/1102) | 84.1% | 98.3% | 302 |
| 5 | van Dommelen L | 2010 | USA | Biorapid CHLAMYDIA Ag test (Biokit, S.A., Barcelona, Spain) | Vaginal swabs | Roche Amplicor CT/NG test | NA | 737 females | 10.2 | 17.1% (13/75) | 93.7% (619/662) | 24.6% | 90.4% | 3 |

| | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| | | | | QuickVue Chlamydia test (Quidel Corporation, San Diego, USA) | Vaginal swabs | Roche Amplicor CT/NG test | NA | 737 females | 9.5 | 25% (14/70) | 99.7% (665/667) | 91.3% | 91.5% | 83 |
| 6 | van der Helm JJ | 2012 | Netherland | Chlamydia Rapid Test (CRT) (Diagnostics for the Real World (Europe), Cambridge, UK) | Vaginal swabs | NAAT (Aptima, Gen-Probe, San Diego, USA) | NA | 912 females | 11.2 | 41.2% (42/102) | 96.4% (781/810) | 59.2% | 92.9% | 19 |
| 7 | Hurly DS | 2014 | Vanuatu | Chlamydia Rapid Test (CRT) | First void male urine samples | COBAS TaqMan 48 PCR assay (Roche Diagnostic, Branchburg, New Jersey, USA) | NA | 156 males | 18.6 | 41.4% (12/29) | 89% (113/127) | 46.2% | 86.9% | 6 |
| | | | | Chlamydia Rapid Test (CRT) | Vaginal swabs | COBAS TaqMan 48 PCR assay (Roche Diagnostic, Branchburg, New Jersey, USA) | NA | 223 females | 27.8 | 74.2% (46/62) | 95.7% (154/161) | 86.8% | 90.6% | 63 |
| | | | | ACON Chlamydia Rapid Test Device, ACON Laboratories (ACON) | Urine sample | COBAS TaqMan 48 PCR assay (Roche Diagnostic, Branchburg, New Jersey, USA) | NA | 132 males | 12.1 | 43.8% (7/16) | 98.3% (114/116) | 77.8% | 92.7% | 44 |

| | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| | | | | ACON Chlamydia Rapid Test Device, ACON Laboratories (ACON) | Vaginal swabs | COBAS TaqMan 48 PCR assay (Roche Diagnostic, Branchburg, New Jersey, USA) | NA | 75 females | 8.0 | 66.7% (4/6) | 91.3% (63/69) | 40.0% | 96.9% | 21 |
| 8 | Nuñez-Forero L | 2016 | Colombia | ACON Plate CT RT | Endocervical swab | COBAS AMPLICOR CT/NG system (Roche) | NA | 229 females | 9.6 | 22.7% (5/22) | 100.0% (207/207) | 100.0% | 92.4% | 130 |
| | | | | QuickVue Chlamydia RT (Quidel Corporation, San Diego, California, USA) | Endocervical swab | COBAS AMPLICOR CT/NG system (Roche) | NA | 664 females | 8.3 | 37.74% (20/53) | 99.4% (607/611) | 83.3% | 94.8% | 92 |
| | | | | CT Duo test combo (ACON Laboratories, San Diego, California, USA) | Endocervical swab | COBAS AMPLICOR CT/NG system (Roche) | NA | 491 females | 12.0 | 30.5% (18/59) | 99.8% (431/432) | 94.7% | 91.3% | 189 |
| 9 | Abbai-Shaik NS | 2016 | South Africa | Chlamydia Rapid Test (CRT) | Urine sample | The BD ProbeTec ET PCR Assay | NA | 100 males | 10.0 | 20.0% (2/10) | 100.0% (90/90) | 100.0% | 91.8% | 53 |
| 10 | Stephen S | 2017 | Zimbabwe | Cortez Onestep Chlamydia RapicardTM insta test | Endocervical swab | Strand Displacement Amplification (SDA) technology | NA | 242 females | 5.8 | 71.4% (10/14) | 99.6% (227/228) | 90.9% | 98.3% | 568 |
| 11 | Saison F | 2007 | Philippines | Clearview Chlamydia MF | Cervical swab | Roche Amplicor CT/NG PCR-based assay | NA | 821 females | 32.0 | 53.5% (85/158) | 99.1% (657/663) | 93.4% | 90.0% | 128 |

| | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| | | | | Clearview Chlamydia MF | Vaginal swab | Roche Amplicor CT/NG PCR-based assay | NA | 333 females | 17.9 | 31.1% (19/61) | 95.2% (259/272) | 59.4% | 86.0% | 9 |
| | | | | Chlamydia Rapid Test (CRT) | Vaginal swab | Roche Amplicor CT/NG PCR-based assay | NA | 291 FSWs | 17.9 | 71.0% (66/93) | 99.0% (196/198) | 97.1% | 87.9% | 240 |
| | | | | Chlamydia Rapid Test (CRT) | Vaginal swabs | Roche Amplicor CT/NG PCR-based assay | NA | 838 females | 6.3 | 86.8% (46/53) | 99.6% (782/785) | 93.9% | 99.1% | 1713 |
| 12 | Wisniewsk i CA | 2008 | UK | DRW Chlamydia Rapid Test | Urine - urine cup | Roche Amplicor CT/NG PCR assay | NA | 534 males | 6.4 | 47.0% (16/34) | 98.8% (494/500) | 72.8% | 96.5% | 73 |
| | | | | DRW Chlamydia Rapid Test | Urine - first burst sample | Roche Amplicor CT/NG PCR assay | NA | 534 males | 6.4 | 82.0% (28/34) | 98.8% (494/500) | 82.4% | 98.8% | 384 |
| 13 | Bandea CI | 2009 | USA | BioStar Chlamydia OIA | Cervical swab | Abbott ligase chain reaction | NA | 261 females | 16.0 | 59.4% (25/42) | 98.4% (216/220) | 87.6% | 92.7% | 79 |
| 14 | Ham JY | 2015 | Korea | aQcare Chlamydia TRF kit (Medisensor, Inc., Daegu, Korea) | Urine | NAATs, the AccuPower CT & NG real-time PCR kit (Bioneer, Inc., Daejeon, Korea) | NA | 93 males and females | 18.3 | 88.2% (15/17) | 94.7% (72/76) | 78.9% | 97.3% | 135 |
| | | | | aQcare Chlamydia TRF kit (Medisensor, Inc., Daegu, Korea) | Swab | NAATs, the AccuPower CT & NG real-time PCR kit (Bioneer, Inc., Daejeon, Korea) | NA | 348 males and females | 27.9 | 93.8% (91/97) | 96.8% (243/251) | 91.9% | 97.6% | 461 |