**Supplementary Table 2**. Network meta-analysis results in the percentage of IOP reduction at follow-up end (Inconsistency Model).

| Weighted mean difference (95% CI), % | | | |
|---|---|---|---|
| 1-iStent | | | |
| 0.62 (-9.89, 9.96) | 2-iStent | | |
| -1.03 (-10.66, 9.24) | -1.62 (-9.84, 8.69) | Hydrus | |
| 7.82 (0.21, 15.43)* | 7.19 (1.27, 14.75)* | 8.80 (2.10, 14.92)* | Control |

*Statistical significance. IOP, intraocular pressure; CI, credible interval; 1-iStent, 1-iStent implantation with phacoemulsification; 2-iStent, 2-iStent implantation with phacoemulsification; Hydrus, Hydrus implantation with phacoemulsification; Control, phacoemulsification alone.