

EEG detector: TCN, seizure vs non-seizure, fixed last epoch (259.4 k parameters)



Receptive field 373 samples (≈ 3 s) inside each 6 s window. Loss: class-weighted CE on window labels (ictal if $\geq 80\%$ of the window lies in the annotated seizure). EEG detects seizures well (macro-F1 ≈ 0.98) but ranks severity at chance within a session, so the model is trained for seizure vs non-seizure only.