GLOBAL CALIBRATION

H.O. : CS-10, UM Floor, Ansal Corporates, Opp. Dabur Chowk, Vaishali, Gzb. (U.P.) Ph.: 0120-4132122

e-mail: support@globalcalibration.com • **Visit us at :** www.globalcalibration.com

CC - 3071

Calibration Certificate

Certificate No.: GC/AV/0530D/10/26	**ULR No.:** CC307124000010328F

Customer Details:
M/s ZTE Telecom India Pvt. Ltd.
Plot No.05, Sector-2A, IMT Manesar 122052
Gurgaon-Haryana.

SRC No.: 0530		**Job No.:** 0530 D	**Pg:** 01/01
Date of Calib.:	03-10-2026	**Sugg. Due Dt.:**	03-10-2027
Date of Issue:	05-10-2026	**Receipt Date:**	03-10-2026
Condition:	OK	**SRC Date:**	07-09-2026

INSTRUMENT DETAILS

Instrument Name: DC Power Supply	**Make:** -	**Model:** DH1715A-5	**Location:** Repair Section
Instrument Sr. No.: 1097335	**I.D:** -	**Range:** 0 to 35V, 0 to 3A DC	**L.C/Resolution:** 0.1V, 0.1A DC
Env. Condition: **Temperature:** 25±4 °C **Humidity:** 30-75% RH	**Calibration Site:** **ON SITE**	**Discipline:** Electro-Technical	**Group:** Direct Current
Accuracy: As per Manual	**Calibration Procedure No./ IS Std.:** GC/WI/002&TE/003, GC/WI/002&TE/09, GC/WI/007		

THE EQUIPMENT USED FOR CALIBRATION IS/ARE TRACEABLE TO NATIONAL/INTERNATIONAL STANDARDS AS PER FOLLOWING DETAILS:

S.No.	Instrument Name	ID / S.No	Traceable to NPL Via Cert. No.	Calib. Due Date
1	8½ Reference Multimeter	489976556 / GC-074	AIPL/171023/1-1	03-10-2027

CALIBRATION RESULTS

Parameter - 1: DC VOLTAGE

S.No	UUC Value (V)	Std./Measured Value (V)	Error (V)	Exp. Uncert. Of Meas (±), k=2	Remarks
1	5.2	4.987513	0.212	0.00003	-
	10.2	9.9824541	0.218	0.00007	-
	20.1	19.897987	0.202	0.00013	-
	30.2	29.987746	0.212	0.00020	-
	35.2	34.947972	0.252	0.00023	-

Parameter - 2: DC Current

S.No	UUC Value (A)	Std./Measured Value (A)	Error (A)	Exp. Uncert. Of Meas (±), k=2	Remarks
2	1.1	0.987841	0.112	0.00017	-
	2.1	1.9856877	0.114	0.00029	-
	3.1	2.9766589	0.123	0.00043	-

Parameter - 3: DC VOLTAGE

S.No	UUC Value (V)	Std./Measured Value (V)	Error (V)	Exp. Uncert. Of Meas (±), k=2	Remarks
1	5.2	4.988961	0.211	0.00003	-
	10.1	9.9875964	0.112	0.00007	-
	20.1	19.976179	0.124	0.00013	-
	30.2	29.988321	0.212	0.00020	-
	35.1	34.988332	0.112	0.00023	-

Prameter - 4: DC Current

S.No	UUC Value (A)	Std./Measured Value (A)	Error (A)	Exp. Uncert. Of Meas (±), k=2	Remarks
2	1.1	0.9987844	0.101	0.00017	-
	2.2	1.9897465	0.210	0.00031	-
	3.1	2.9898197	0.110	0.00043	-

Remarks

1. UUC - Indicates 'Unit Under Calibration'. This Certificate refers only to the particular item submitted for the calibration. Calibration is performed for Scientific and research work only.
2. The uncertainty of measurement at 95% (k=2)confidence level is as indicated in cal. certificate. The results reported in this Certificate are Valid at the time of and under the stipulated conditions of measurement.
3. The Reported values are average of five set of reading. The Liability of calibration is limited/ equivalent to the calibration cost of instrumentpaid by the company Director of Global Calibration.
4. This Certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Director of Global Calibration.

Calibrated By	*** End of Calibration ***	Authorised By
		J. Malhotra