

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



log no - 2085

شركة اتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 58264

Customer	Aloof Hospital Jahra		
Equipment	OT Light		
Model No.	Penta 63N		
Serial No.	029578		
Authorized Person	Manufacturer RIMSA		
Customer Complaint	Location OR		
	Contract / Order No.		
	☐ Installation ☒ Warranty		
	☒ Repair ☐ On Charge		
	☐ Maintenance ☐ Contract		
	☐ Upgrade ☐ _____		
	Time Spent		
Services / Check Carried out			
☐ Inspection / Cleaning	☐ Optics Checks		
☐ Calibration / Performance	☐ Pneumatic Checks		
☐ Interface / Communication	☐ Hydraulic Checks		
☐ Safety Checks	☐ Operational Error		
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		
Remarks	Replaced the power base with new. Checked all the function and found it working properly		
No.	Spare Parts Used	Part No.	Qty.
	replaced		
	new		
	warranty		
Service Engineer	Authorized Person	Customer / User	
Signature	Signature	Signature	
Date	Date	Date	
11/3/24	11/2/24		