

INCIDENT REPORT FORM

LOG #:	Forward Completed Form to Department Manager. One copy to Quality Management Department.		
PART A. Status: ☐ Inpatient ☐ Outpatient ☐ Visitor/Family ☒ Employee ☐ Other (specify) _____			
Date of Occurrence: <u>08-04-25</u>	Time of Occurrence: <u>5 PM</u>	Occurrence Location: <u>5TH FLOOR ALORF HOSPITAL</u>	
If event involved employee - Staff Name: <u>ROSENA ERIKA CABILES</u>	Age: <u>37yrs</u> Sex: <u>F</u> Emp#: <u>2844</u>		
If event involved patient - Patient Name: _____	Age: _____ Sex: _____ File#: _____		
If event involved Visitor - Visitor Name: _____	Phone No.: _____		
Address: _____			

PART B. Occurrence Type (check boxes that apply)

INTRAVENOUS RELATED	PROCEDURE VARIANCE	HAZARD/SAFETY
☐ Central Line ☐ Contrast Media Reaction ☐ Extravasate ☐ Occluded ☐ Precipitate ☐ Site Red/Swollen ☐ Tubing Leak ☐ Expired IV Fluid ☐ Additive Related ☐ Transfusion Reaction ☐ Other _____	☐ Delays ☐ Missing Files, Notes ☐ Performed on Wrong Site ☐ Consent ☐ Patient Identification ☐ Wrong Patient ☐ Wrong Procedure ☐ Lost/Spoiled Specimen ☐ No Response ☐ Wrong File ☐ Procedure Cancellation ☐ Notes in Wrong File ☐ Inadequate/Inappropriate Entry ☐ No label/Addressograph ☐ Other _____	☐ Burn ☐ Chemical Spill ☐ Electrical Shock ☐ Fire ☐ Gas ☐ Hazardous Waste ☐ Radiation Exposure ☐ Splash ☐ Water Spill ☐ Needle prick injury ☐ Other _____
BEHAVIOR	SECURITY VARIANCE	EQUIPMENT/SUPPLIES
☐ Improper disposal of sharp ☐ AMA/Absconded ☐ Dissatisfied with care ☐ Patient Behavior ☐ Family/Visitor Behavior ☐ Staff Behavior ☐ Time In/Time Out Variation ☐ Non-Compliance ☐ Patient Complaint ☐ Other _____	☐ Damage/Loss of Property ☐ Kidnapping ☐ Missing/Misplaced Drug Keys ☐ Prescription Alteration ☐ Security Problem ☐ Suicide attempt ☐ Other _____	☐ Electrical Problem ☒ Not Working ☐ Improper Use ☒ Mechanical Problem ☐ Operator Not Qualified ☐ Wrong Equipment ☐ Missing Equipment ☐ Not Available ☐ Out of Stock ☐ Other <u>ENDOSCOPE & ENT WORK STATION</u>

ADVERSE DRUG REACTION (ADR) / MEDICATION OCCURENCE

☐ Medication missing	☐ Wrong Patient	Describe Therapy:
☐ Adverse Side effects	☐ Others:	
☐ Medication given but not charted		
☐ Medication charted but not given	*please see Intravenous Related area	
☐ Duplication/extra dose given	Describe reaction (chronological sequence):	
☐ Time variance (wrong time)		Response to Treatment/Outcome:
☐ Wrong route		
☐ Wrong dose		

Reported by	Name: _____	ID No. _____	Dept. _____	Position _____
IMMEDIATE ACTION Physician Notified ☐ Yes ☐ No. Did MD see patient ☐ Yes ☐ No ☐ X-ray ☐ Suture ☐ Blood Work ☐ Medication				

Assessment/ Diagnosis			
Physician Name:	Dept.:	ID No.:	

INCIDENT REPORT FORM

PART C. Describe the occurrence (state the facts). Use when necessary.

FACTUAL DESCRIPTION

- ① ENDOSCOPE MACHINE MONITOR IS NOT WORKING - NO DISPLAY
- ② ENT WORK STATION - WATER LEAK COMING FROM WASH TIP HOSE WHEN USED

PART D. Severity Codes (Please check one).

- ☐ Minor. No injury or minor treatment, i.e. abrasion, No increased level of care.
- ☐ Moderate. Increased length of Stay. Increased level of care.
- ☐ Major. Temporary/permanent lessening of bodily functions (sensory, motor, physiologic, or intellectual). Not related to the natural Outcome of the patient's illness, underlying condition, or disfigurement. Surgical intervention required.
- ☐ Sentinel/Catastrophic. Death/Major permanent loss of function (sensory, motor, physiologic, or intellectual), hemolytic transfusion reaction, surgery on wrong patient or body part.
- ☐ Near Miss. For potentially severe Outcome.

PART E. ACTIONS.

PREVENTIVE ACTION

ACTION TAKEN BY INVOLVED DEPARTMENT		REFERRED TO QUALITY MANAGEMENT DEPT.	
referred to Biomed DEPT. they will arrange to be seen by the engineer.		- Referred to Biomedical team. "Root Cause Analysis" to put recommendation & prevent its recurrence.	
Dept. Head: <u>CHRISTINA CUEVA</u> Date: <u>8/4/2024</u>		Quality Mgmt. Dept. <u>Charlotte TB / RN</u> Date: <u> </u>	
ID No. <u>1022</u> Department: <u>NURSING</u>		Comments: <u> </u>	

Department Managers: Forward Completed forms to Quality Management Department

MEDICAL EQUIPMENT REPAIR REQUEST

No. _____

Date: 08-04-25

Requested by ROWENA ERIKA CABILES

Department ENT Location 5TH FLOOR ALORF HOSPITAL

SR. No.	Machine	Model	Complaint
① <u>YN2000597</u>	<u>AIDA C</u>	<u>WD 310</u>	<u>- MONITOR (NO DISPLAY)</u> <u>- LOOSE CABLE CONNECTION</u>
② <u>111692045</u>	<u>ENT WORK STATION</u>	<u>ENTERMED</u>	<u>- WATER LEAK</u>

Check within

☐ Week

☒ ASAP

Engineer Comments

End Comments


ROWENA ERIKA CABILES

Requested by

Checked By

Approved by

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٥٦٧٧٢٥ / ٢٤٧٤٣٠١٧ (+٩٦٥)

فاكس : ٢٤٥٦٧٧٩٢ / ٢٤٧٤٣٠١٩ (+٩٦٥)

ص.ب : ٢٣٥٥ الجبراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 65371

Customer	Alert Hospital	Equipment	ENT work station
		Model No.	Futurent III
		Serial No.	111632045
Authorized Person		Manufacturer	Entermed.
Customer Complaint		Location	5th floor.
		Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☒ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ _____
☐ Safety Checks	☐ Operational Error	Time Spent	
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		
Remarks	Service call attended. we inspect the unit and found that the water leakage from the irrigation handpiece. The irrigation handpiece thread is damaged. Need to replace the irrigation hand piece. Quotation will be submitted as soon as possible.		
No.	Spare Parts Used	Part No.	Qty.
Service Engineer	Authorized Person	Customer / User	
Mixon			
ATCO Technology Co.			
Signature	Signature	Signature	
Date 09/04/2025	Date	Date 10-4-25	