

1493

Roche for the Trade in Medicines, Equipment, Devices
and Medical Supplies SPC
Customer Report Form



Case number CAS-0020245532

Order number WO-28669999 | Visit Date 09/27/2024

Instrument Details		Customer Details	
Instrument/Module	cobas b 221<6>=Roche OMNI S6 system	Customer Number	0050250925
Serial Number	19239	Customer Name	Dr. Mansour Alorf Hospital
Additional Details	PO	Street Address	Mansour Saad Al Bannaq Street
		Zip	01025 - Jahra
		Contact Name	Gina Bisenio
		Contact Phone	+96565991613

Service Activity Code Service Visit

Purpose of Visit

CUSTOMER ALLEGATION

=====

Created by: Gina Bisenio, biomed@alorf.com (+96565991613)

Preferred callback time:

Instrument: b 221 6 Roche OMNI S6 system (Blood Gas/Electrolytes) (Serial: 19239)

Alarm code: not-applicable

Issue type: OTHER

Lot no/Vendor: not applicable

Initial description: ph,po2,pco2,ise error in screen

Issue timeline:

1: Issue created (02.05.2024 15:55 GMT+00:00)

ACTIONS

=====

Type your descriptive analysis here

CONCLUSION

=====

Log a call or note

Performed Activities

Checked FMS tubing set

Unit is okay

Time Report

Category	Start	End	Hours
05913616001-Service labour time	05/24/2024 9:55 pm	05/24/2024 10:55 pm	1.00

Signature or Stamp

Customer
Gina Bisenio

Roche Support
Mohamed Elbekey

Sign Here

Handwritten signature of Mohamed Elbekey over a background of faint, repeating text.

By signing this document

- Roche Support person confirms the above mentioned work and checks have been completed..
- Customer representative confirms they have reviewed the work carried out as described and accepts the system for use.
- Thank you for your trust.

Address: Al Mirqab, Al Soor Street, Block 1, Burj Al Shaya, Plot 6, Floor 5, Office
11. P.O. Box: 467, Safat - 13005 Kuwait
Hotline: 25770360
Fax: 22059262

Customer Assistance: