

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجهر، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63008

Customer	Alorif Hospital			Equipment	Autoclave		
				Model No.	Selectomat PL		
				Serial No.	220823		
Authorized Person				Manufacturer	MMM		
Customer Complaint				Location	CSSA		
No Vacuum in Chamber				Contract / Order No.			
Services / Check Carried out ☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check ☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐				☐ Installation		☐ Warranty	
				☒ Repair		☐ On Charge	
				☐ Maintenance		☐ Contract	
				☐ Upgrade		☐	
				Time Spent			
Remarks we found that the two valves are faulty. we replaced the two valves with new ones. we tested the unit with vacuum test, BBW test and cycle. checked and confirmed that the unit is working properly.							
No.	Spare Parts Used			Part No.		Qty.	
Service Engineer		Authorized Person			Customer / User		
Signature		Signature			Signature		
Date		Date			Date		
24/04/2025		20/5/25					