

FIELD SERVICE REPORT

LCS 12905

[illegible]

PROBLEM, DIAGNOSIS AND ACTION

- ④ Service call attended
- ④ unit checked
- ④ problem: Air Probe failure. The Temperature of the machine was not increasing now it has been fixed and temp. increased
- ④ unit is in good condition

REMARKS					
JOB COMPLETED	☒ Yes	☐ No	AWAITING PARTS	☐ Yes	☒ No
ATC ENGINEER SIGNATURE		USER SIGNATURE		AUTHORIZED SIGNATURE & STAMP	
					