

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجهرام, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 61018

Customer	Alf Hospital Jahra	Equipment	G-Runner Interface
		Model No.	96368
		Serial No.	24000771
Authorized Person		Manufacturer	FOMDA
Customer Complaint		Location	1st FLR - OB
		Contract / Order No.	

Services / Check Carried out

- | | |
|---|---|
| ☐ Inspection / Cleaning | ☐ Optics Checks |
| ☐ Calibration / Performance | ☐ Pneumatic Checks |
| ☐ Interface / Communication | ☐ Hydraulic Checks |
| ☐ Safety Checks | ☐ Operational Error |
| ☐ Electrical / Electronics | ☐ Mechanical Checks |
| ☐ Accessories Check | ☐ |

- | | |
|---|--|
| ☐ Installation | ☐ Warranty |
| ☒ Repair | ☒ On Charge |
| ☐ Maintenance | ☐ Contract |
| ☐ Upgrade | ☐ |

Time Spent

Remarks

Replaced the interface module for the G-Runner with new. Checked all the functions and found it working in good condition.

For SP Dynamics - SN: 19101628

No.	Spare Parts Used	Part No.	Qty.
①	G-Runner Interface		1

Service Engineer

ATCO Technology Co.

Signature

Date

Signature
Date 16/10/2024

Authorized Person

Signature

Date

Signature
Date 14/10/2024

Customer / User

Signature

Date