

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجھراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 58338

Customer	Alorf Hospital.		
Equipment	OBI Scanner,		
Model No.			
Serial No.	009425 B8		
Authorized Person	Manufacturer OBIScanner,		
Customer Complaint	Location Clinical Dr Asem		
OB1 Scanner license expired	Contract / Order No.		
Services / Check Carried out	☐ Installation ☐ Warranty		
☐ Inspection / Cleaning	☐ Optics Checks	☒ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ _____
☐ Safety Checks	☐ Operational Error	Time Spent	
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		
Remarks	We checked the unit and validate the software license with the help of manufacture through online supports now the unit is working properly.		
No.	Spare Parts Used	Part No.	Qty.
Service Engineer	Authorized Person	Customer / User	
Miscan		د. ناصر محمد كامل	
ATCO Technology Co.		اختصاصي علاج الشفة	
Signature	Signature	Signature	
Date 28/02/2024	Date	Date 28-02-2024	