

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٧٧٢٥ / ٢٤٧٤٣٠١٧ (+٩٦٥)

فاكس : ٢٤٥٦٧٧٩٢ / ٢٤٧٤٣٠١٩ (+٩٦٥)

ص.ب : ٢٣٥٥ الجھراء، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 56965

Customer	Al-Orf Hospital Zuhrah.	Equipment	Dental Chair
		Model No.	
		Serial No.	Reptom - 6.
Authorized Person		Manufacturer	
Customer Complaint		Location	
	Air & Water Connections from Dental Chair for Scalor.	Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☐ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ _____
☐ Safety Checks	☐ Operational Error		
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		
		Time Spent	

Remarks

Installed Air & Water line connections
for Scalor unit - from Dental
Chair line.
Quick connection placed outside chair
for easy connection & disconnection

No.	Spare Parts Used	Part No.	Qty.

Service Engineer Ahmed. K. Al-Murjan ATCO Technology Co. Signature Date 25-4-24	Authorized Person Signature Date	Customer / User Amr. K. Al-Murjan طبيب مسجل الاستشارة الصناعية مستشفى السورف 22024100048 Signature Date
---	--	---