

INCIDENT REPORT FORM

LOG #:	Forward Completed Form to Department Manager. One copy to Quality Management Department.		
PART A. Status: ☐ Inpatient ☒ Outpatient ☐ Visitor/Family ☐ Employee ☐ Other (specify) _____			
Date of Occurrence: <u>06-07-25</u>	Time of Occurrence: <u>1845H</u>	Occurrence Location: <u>5th FIK ENT</u>	
If event involved employee - Staff Name: <u>ROWENA ELIEN CABIBS</u>		Age: <u>37y</u>	Sex: <u>F</u> Emp#: <u>2844</u>
If event involved patient - Patient Name: _____		Age: _____	Sex: _____ File#: _____
If event involved Visitor - Visitor Name: _____		Phone No.: _____	
Address: _____			

PART B. Occurrence Type (check boxes that apply)

INTRAVENOUS RELATED	PROCEDURE VARIANCE	HAZARD/SAFETY
☐ Central Line ☐ Contrast Media Reaction ☐ Extravasate ☐ Occluded ☐ Precipitate ☐ Site Red/Swollen ☐ Tubing Leak ☐ Expired IV Fluid ☐ Additive Related ☐ Transfusion Reaction ☐ Other _____	☐ Delays ☐ Missing Files, Notes ☐ Performed on Wrong Site ☐ Consent ☐ Patient Identification ☐ Wrong Patient ☐ Wrong Procedure ☐ Lost/Spoiled Specimen ☐ No Response ☐ Wrong File ☐ Procedure Cancellation ☐ Notes in Wrong File ☐ Inadequate/Inappropriate Entry ☐ No label/Addressograph ☐ Other _____	☐ Burn ☐ Chemical Spill ☐ Electrical Shock ☐ Fire ☐ Gas ☐ Hazardous Waste ☐ Radiation Exposure ☐ Splash ☐ Water Spill ☐ Needle prick injury ☐ Other _____
BEHAVIOR	SECURITY VARIANCE	EQUIPMENT/SUPPLIES
☐ Improper disposal of sharp ☐ AMA/Absconded ☐ Dissatisfied with care ☐ Patient Behavior ☐ Family/Visitor Behavior ☐ Staff Behavior ☐ Time In/Time Out Variation ☐ Non-Compliance ☐ Patient Complaint ☐ Other _____	☐ Damage/Loss of Property ☐ Kidnapping ☐ Missing/Misplaced Drug Keys ☐ Prescription Alteration ☐ Security Problem ☐ Suicide attempt ☐ Other _____	☐ Electrical Problem ☐ Not Working ☐ Improper Use ☒ Mechanical Problem ☐ Operator Not Qualified ☐ Wrong Equipment ☐ Missing Equipment ☐ Not Available ☐ Out of Stock ☐ Other <u>water leaking from work-station / EM machine</u>

ADVERSE DRUG REACTION (ADR) / MEDICATION OCCURENCE

☐ Medication missing	☐ Wrong Patient	Describe Therapy:
☐ Adverse Side effects	☐ Others:	
☐ Medication given but not charted		
☐ Medication charted but not given	*please see Intravenous Related area	
☐ Duplication/extra dose given	Describe reaction (chronological sequence):	
☐ Time variance (wrong time)	Response to Treatment/Outcome:	
☐ Wrong route		
☐ Wrong dose		

Reported by	Name: _____	ID No. _____	Dept. _____	Position _____
IMMEDIATE ACTION Physician Notified ☐ Yes ☐ No. Did MD see patient ☐ Yes ☐ No ☐ X-ray ☐ Suture ☐ Blood Work ☐ Medication				

Assessment/ Diagnosis

Physician Name: _____	Dept.: _____	ID No.: _____

INCIDENT REPORT FORM

PART C. Describe the occurrence (state the facts). Use when necessary.

FACTUAL DESCRIPTION

AFTER DOING EAR WASH PROCEDURE, WATER HAS BEEN LEAKING EXCESSIVELY. THE MAIN FILTERS HAVE BEEN CHANGED RECENTLY. NO VISIBLE BLOCKAGES WERE SEEN. WE DECIDED TO TURN THE MACHINE OFF ENTIRELY FOR THE PATIENT & STAFF SAFETY. ALL ELECTRICAL OUTLETS LOCATED NEAR THE MACHINE HAVE ALSO BEEN TURNED OFF. A REPAIR REQUEST HAS BEEN SUBMITTED AND ARE JUST AWAITING FURTHER UPDATE REGARDING THIS.

PART D. Severity Codes (Please check one).

- ☐ **Minor.** No injury or minor treatment, i.e. abrasion, No increased level of care.
- ☐ **Moderate.** Increased length of Stay. Increased level of care.
- ☐ **Major.** Temporary/permanent lessening of bodily functions (sensory, motor, physiologic, or intellectual). Not related to the natural Outcome of the patient's illness, underlying condition, or disfigurement. Surgical intervention required.
- ☐ **Sentinel/Catastrophic.** Death/Major permanent loss of function (sensory, motor, physiologic, or intellectual), hemolytic transfusion reaction, surgery on wrong patient or body part.
- ☐ **Near Miss.** For potentially severe Outcome.

PART E. ACTIONS.

PREVENTIVE ACTION

ACTION TAKEN BY INVOLVED DEPARTMENT		REFERRED TO QUALITY MANAGEMENT DEPT.	
7 referred to Biomed		Equipment checked and referred to Biomedical to identify source and resolve issue.	
Dept. Head: CHRISTINA CUREVA Date: 6/7/2016		Quality Mgmt. Dept. Date:	
ID No. 1622 Department: Nursing		Comments:	

Department Managers: Forward Completed forms to Quality Management Department

MEDICAL EQUIPMENT REPAIR REQUEST

No. _____

Date: 06-07-2025

Requested by ROWENA PRIMA CABILES

Department ENT Location 5th Floor ENT (Dr. Krishna's Clinic)

SR. No.	Machine	Model	Complaint
111632045	ENT Work Station	Futurment III	EXCESSIVE WATER LEAK

Check within Week

☒ ASAP

Engineer Comments

End Comments


ROWENA PRIMA CABILES

Requested by

Checked By

Approved by

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكوتكنولوجيه

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب. ٢٣٥٥ الجراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63794

Customer	AL RO. ALDRF Hospital. Jahra.	Equipment	ENT Workstation
		Model No.	Futurcent 3.
		Serial No.	111632045.
Authorized Person		Manufacturer	Enteamed.
Customer Complaint	Water Leaking.	Location	ENT - 5 th Fl.
		Contract / Order No.	
Services / Check Carried out	☐ Installation ☐ Warranty ☐ Repair ☐ On Charge ☐ Maintenance ☐ Contract ☐ Upgrade ☐ _____		
☐ Inspection / Cleaning ☐ Optics Checks ☐ Calibration / Performance ☐ Pneumatic Checks ☐ Interface / Communication ☐ Hydraulic Checks ☐ Safety Checks ☐ Operational Error ☐ Electrical / Electronics ☐ Mechanical Checks ☐ Accessories Check ☐ _____	Time Spent		
Remarks	Checked and found drainage line is blocked. Open the drain passages and cleaned. Checked and After cleaned the system, there is no leak found. The Unit is working properly.		
No.	Spare Parts Used	Part No.	Qty.
Service Engineer	Authorized Person	Customer / User	
CHERIAN / Jince.			
ATCO Technology Co.			
Signature	Signature	Signature	
Date	Date	Date	
07/07/2025	8/7/25		