

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجھراء , ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63782

Customer	Alorf Hospital			Equipment	Patient bed.		
				Model No.	LA 6018 MAPL		
				Serial No.	058461		
Authorized Person				Manufacturer	Corona		
Customer Complaint	wheel brake faulty.			Location	HA AL ER.		
				Contract / Order No.			
Services / Check Carried out	☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check			☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐	☐ Installation ☒ Repair ☐ Maintenance ☐ Upgrade	☐ Warranty ☐ On Charge ☐ Contract ☐	Time Spent
Remarks	we repaired the Patient bed brake system. we removed the 4 wheels and cleared the issue and replaced. checked and confirmed that the unit is working properly						
No.	Spare Parts Used			Part No.	Qty.		
Service Engineer	Authorized Person			Customer / User			
Nixon/Manu ATCO Technology Co.	for 10/12/2024						
Signature	Signature			Signature			
Date 8/12/2024	Date			Date			