

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجهر، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 59407

Customer	Alom Hospital.			Equipment	Shock wave therapy unit.		
				Model No.	SHOCK MED		
				Serial No.	EM09311018		
Authorized Person				Manufacturer	EME		
Customer Complaint				Location	physiotherapy		
				Contract / Order No.			
Services / Check Carried out ☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check ☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐				☐ Installation	☐ Warranty		
				☒ Repair	☐ On Charge		
				☐ Maintenance	☐ Contract		
				☐ Upgrade	☐		
				Time Spent			
Remarks							
Service call attended. we inspect the unit and found that the touch screen not working properly. The sensitivity for the touch screen not perfectly working. we will check with the manufacturer and get back ASAP.							
No.	Spare Parts Used			Part No.	Qty.		
Service Engineer		Authorized Person		Customer / User			
Mishal		for 16/04/2023					
Signature		Signature		Signature			
Date 06/04/2023		Date		Date			