

ATCO Technology Co.

Tel : (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 60269

Customer	Alor Hospital			Equipment	Audiometer		
				Model No.	Vida		
				Serial No.	IMDD17203836		
Authorized Person				Manufacturer	Inventis		
Customer Complaint				Location	ENT		
not working.				Contract / Order No.			
Services / Check Carried out				☐ Installation			☐ Warranty
				☒ Repair			☐ On Charge
☐ Inspection / Cleaning				☐ Optics Checks			
☐ Calibration / Performance				☐ Pneumatic Checks			
☐ Interface / Communication				☐ Hydraulic Checks			
☐ Safety Checks				☐ Operational Error			
☐ Electrical / Electronics				☐ Mechanical Checks			
☐ Accessories Check				☐ _____			
				Time Spent			
Remarks							
Service call attended.							
we inspect the unit and found the connections are not properly done. we reconnect all modalities and checked the audiometer.							
found that the unit is working properly.							
checked and confirmed that the unit is working properly.							
No.	Spare Parts Used			Part No.	Qty.		
Service Engineer		Authorized Person		Customer / User			
Nixon/ Jeyapalan		for 16/12/2021					
ATCO Technology Co.							
Signature		Signature		Signature			
Date 01/06/2022		Date		Date			