

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجبراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63835

Customer	Aloft Hospital	Equipment	Electro surgical unit
		Model No.	Surgilase MC3
		Serial No.	208
Authorized Person		Manufacturer	Q. Surgilase
Customer Complaint		Location	OR
		Contract / Order No.	

Services / Check Carried out	☐ Installation	☐ Warranty
	☒ Repair	☐ On Charge
☐ Inspection / Cleaning	☐ Optics Checks	☐ Maintenance
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade
☐ Safety Checks	☐ Operational Error	☐
☐ Electrical / Electronics	☐ Mechanical Checks	
☐ Accessories Check	☐	
	Time Spent	

Remarks	Service call attended
	We inspect the unit and found that the
	Neutral connection is not detected.
	We checked the connection, need to replace the connector
	and need to check the board.
	Currently unit is not working

No.	Spare Parts Used	Part No.	Qty.
	still		
	NOT WORKING		
	For repair		

Service Engineer	Authorized Person	Customer / User
N. Joon	20/5/25	
ATCO Technology Co.		
Signature	Signature	Signature
Date	Date	Date
20/02/2025		