

Customer <u>Alor Hospital</u>		Equipment <u>Autoclave</u>	
		Model No. <u>Selectomab PL</u>	
		Serial No. <u>220824</u>	
Authorized Person		Manufacturer <u>MMM</u>	
Customer Complaint		Location <u>CSSD</u>	
		Contract / Order No.	
<u>Stain on cove ring.</u>		☐ Installation ☐ Warranty	
		☒ Repair ☐ On Charge	
		☐ Maintenance ☐ Contract	
		☐ Upgrade ☐ _____	
Services / Check Carried out		Time Spent	
☐ Inspection / Cleaning ☐ Optics Checks			
☐ Calibration / Performance ☐ Pneumatic Checks			
☐ Interface / Communication ☐ Hydraulic Checks			
☐ Safety Checks ☐ Operational Error			
☐ Electrical / Electronics ☐ Mechanical Checks			
☐ Accessories Check ☐ _____			
Remarks <u>Service call attended</u>			
<u>We inspect the unit and found that Stain occurs</u>			
<u>and after the procedure.</u>			
<u>We replaced the drain trap and checked valve.</u>			
<u>we tried some cycle and found that there is no</u>			
<u>Stain occurs. We kept the unit under observation</u>			
No.	Spare Parts Used	Part No.	Qty.
Service Engineer		Authorized Person	
<u>Nixon</u>		<u>11/9/24</u>	
ATCO Technology Co.			
Signature <u>[Signature]</u>		Signature	
Date <u>11/09/2024</u>		Date	
		Customer / User	
		Signature	
		Date	