

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكوتكنولوجيه

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهرء , ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63839

Customer	Alorof Hospital	Equipment	Digital Video Colposcope
		Model No.	AL-106
		Serial No.	121 ADR0033
Authorized Person		Manufacturer	Med gym
Customer Complaint		Location	OK
	Image not clear.	Contract / Order No.	

Services / Check Carried out

- | | |
|--|--|
| ☐ Inspection / Cleaning | ☐ Optics Checks |
| ☐ Calibration / Performance | ☐ Pneumatic Checks |
| ☐ Interface / Communication | ☐ Hydraulic Checks |
| ☐ Safety Checks | ☐ Operational Error |
| ☐ Electrical / Electronics | ☐ Mechanical Checks |
| ☐ Accessories Check | ☐ |

- | | |
|---------------------------------------|------------------------------------|
| ☐ Installation | ☐ Warranty |
| ☐ Repair | ☐ On Charge |
| ☐ Maintenance | ☐ Contract |
| ☐ Upgrade | ☐ |

Time Spent

Remarks: Replaced camera head (SN: 118ABDR0033) with new one (SN: 118AAER0017). Tested and confirmed unit is working OK.

Demo given to staff.

No.	Spare Parts Used	Part No.	Qty.
	New Camera Head	SN: 118AAER0017	

Demo done, working now

Service Engineer

(Signature)
ATCO Technology Co.

Signature

Date

(Signature)
5/3/25

Authorized Person

Signature

Date

(Signature)
17/3/25

Customer / User

Signature

Date