

Tel.: (+965) 24743017 / 24567725
Fax: (+965) 24743019 / 24567792
P.O. Box: 2355 Jahra, 01025 - Kuwait
Email: sales@atco-kuwait.com
www.atco-kuwait.com
CR NO. 336218



شركة أتكو التكنولوجيه

تلفون : ٢٤٥٦٧٧٢٥ / ٢٤٧٤٣٠١٧ (+٩٦٥)

فاکس: ۲۴۵۶۷۷۹۲ / ۲۴۷۴۳۰۱۹ (+۹۶۵)

ص.ب : ٢٣٥٥ الجبراء، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

سن.ت. ۳۳۶۲۱۸

No.: 64144

SERVICE / INSTALLATION REPORT			
Customer Ator Hospital Dahra		Equipment Endoscopy Washer & Disinfectant	
		Model No. Series 3	
		Serial No. 25186	
Authorized Person		Manufacturer Soluscope	
Customer Complaint Error 12-30 Solution 'A'		Location ENDOSCOPY 2ND FLR.	
		Contract / Order No.	
Services / Check Carried out ☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check ☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐ _____		☐ Installation ☐ Warranty	
		☐ Repair ☐ On Charge	
		☐ Maintenance ☐ Contract	
		☐ Upgrade ☐ _____	
		Time Spent	
Remarks Found the chemistry 'A' pump defective, need a replacement and some other solenoid valves are getting defective. Chemistry 'A' pump changed with standby. Done cycle 6 and completed. will send quotation for service and spares.			
No.	Spare Parts Used	Part No.	Qty.
Service Engineer ATCO Technology Co. Signature Date 11/02/2025		Authorized Person Signature Date	
		Customer / User Signature Date	