

ATCO Technology Co.

Tel: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



log no - 2075

شركة اتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 58380

Customer	Alcost Hospital			Equipment	CO ₂ Incubator		
				Model No.	C60		
				Serial No.	4093-0301		
Authorized Person				Manufacturer	Labotect		
Customer Complaint				Location	IVF.		
				Contract / Order No.			
				☐ Installation	☐ Warranty		
				☐ Repair	☐ On Charge		
				☐ Maintenance	☐ Contract		
				☐ Upgrade	☐ EM		
				Time Spent			
Services / Check Carried out							
☐ Inspection / Cleaning			☐ Optics Checks				
☐ Calibration / Performance			☐ Pneumatic Checks				
☐ Interface / Communication			☐ Hydraulic Checks				
☐ Safety Checks			☐ Operational Error				
☐ Electrical / Electronics			☐ Mechanical Checks				
☐ Accessories Check			☐				
Remarks							
Replaced faulty CO ₂ sensors, Calibrated CO ₂ at 6% and O ₂ at 21%. Tested temperature and confirmed unit is working properly.							
No.	Spare Parts Used			Part No.		Qty.	
	CO ₂ sensor					1	
Service Engineer		Authorized Person			Customer / User		
Signature		Signature			Signature		
Date		Date			Date		
1/2/24		MIRA / 1-2-24					