

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



log no - 2588

شركة اتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 57975

Customer	Aloof Hospital		
Equipment	Washer Disinfectant		
Model No.	Uniclean		
Serial No.	230506		
Authorized Person	MMM		
Customer Complaint	Location		
Contract / Order No.	C88D		
Services / Check Carried out	☐ Installation ☐ Warranty		
☐ Inspection / Cleaning	☐ Optics Checks	☐ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐
☐ Safety Checks	☐ Operational Error	Time Spent	
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐		
Remarks	Service call attended. Found unloading door time exceed error. Fixed the problem. Tested with full cycle, unit is working ok.		
No.	Spare Parts Used	Part No.	Qty.
Service Engineer	Authorized Person	Customer / User	
Signature	Signature	Signature	
Date	Date	Date	