

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: **65149**

Customer	Alorf Hospital			Equipment	Laser machine		
				Model No.	SP Dynamics		
				Serial No.	16100425		
Authorized Person				Manufacturer	Fotona		
Customer Complaint				Location	4th floor		
Error H2.				Contract / Order No.			
Services / Check Carried out ☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check ☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐				☐ Installation	☐ Warranty		
				☒ Repair	☐ On Charge		
				☐ Maintenance	☐ Contract		
				☐ Upgrade	☐		
				Time Spent			
Remarks Service call attended. we inspect the unit and found that the Error H2. suspecting the issue with the charger board unit. After fixing the charger board only we can check other internal parts. Once we rectify the issue we will update as soon as possible.							
No.	Spare Parts Used			Part No.	Qty.		
Service Engineer M. Alon ATCO Technology Co. Signature Date 18/03/2025		Authorized Person Signature Date		Customer / User د. ايمن نواف استشاري امراض جلدية وتناسية مستشفى النور 140945 Signature Date			