

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



log no - 2589

شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجهراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 60111

Customer	Aloof Hospital			Equipment	Drying cabinet		
				Model No.	DC 21		
				Serial No.	7295		
Authorized Person				Manufacturer	Malmet		
Customer Complaint				Location	CSSD		
				Contract / Order No.			
				☐ Installation	☒ Warranty		
				☒ Repair	☐ On Charge		
				☐ Maintenance	☐ Contract		
				☐ Upgrade	☐ _____		
				Time Spent			
Services / Check Carried out	☐ Inspection / Cleaning						☐ Optics Checks
	☐ Calibration / Performance						☐ Pneumatic Checks
	☐ Interface / Communication						☐ Hydraulic Checks
	☐ Safety Checks						☐ Operational Error
	☐ Electrical / Electronics						☐ Mechanical Checks
	☐ Accessories Check						☐ _____
Remarks	Service call attended.						
	we inspect the unit and found that the thermostat is faulty. we replaced the thermostat with new one.						
	we kept the unit under observation for one day. checked and confirmed that the unit is working properly.						
No.	Spare Parts Used		Part No.		Qty.		
1	Thermostat DM TRS/1211				1		
Service Engineer	Authorized Person		Customer / User				
Signature	Signature		Signature				
Date	Date		Date				