

1493

Roche for the Trade in Medicines, Equipment, Devices
and Medical Supplies SPC
Customer Report Form



Case number CAS-0019848956

Order number WO-27539393 | Visit Date 03/10/2024

Instrument Details

Instrument/Module cobas b 221<6>=Roche OMNI
S6 system

Serial Number 19239

Customer Details

Customer Number 0050250925

Customer Name Dr. Mansour Alorf Hospital

Street Address Mansour Saad Al Bannaq
Street

Zip 01025 - Jahra

Contact Name Gina Bisenio

Contact Phone +96565991613

Additional Details

PO

Service Activity Code Service Visit

Purpose of Visit

CUSTOMER ALLEGATION

=====

Created by: Gina Bisenio, biomed@alorf.com (+96565991613)

Preferred callback time:

Instrument: b 221 6 Roche OMNI S6 system (Blood Gas/Electrolytes) (Serial: 19239)

Alarm code: heating error

Issue type: OTHER

Lot no/Vendor: not applicable

Issue timeline:

1: Issue created (10.03.2024 8:19 GMT+00:00)

ACTIONS

=====

Type your descriptive analysis here

CONCLUSION

=====

Log a call or note

Performed Activities

After checked found BG SD cartridge fan was not moving

Cleaned all the parts required

Cleaned and lubricant SD fan found BG SD heating was crossed

Run calibration and QC

Performed all the service required

Reset the instrument

All the functions working

Time Report

Category	Start	End	Hours
05913616001-Service labour time	03/10/2024 7:00 pm	03/10/2024 8:30 pm	1.50

Customer
Gina Bisenio

Roche Support
Hazem Ahmed

Sign Here

[illegible]

- Roche Support person confirms the above mentioned work and checks have been completed..
- Customer representative confirms they have reviewed the work carried out as described and accepts the system for use.
- Thank you for your trust.

Address: Al Mirqab, Al Soor Street, Block 1, Burj Al Shaya, Plot 6, Floor 5, Office
11. P.O. Box: 467, Safat - 13005 Kuwait
Hotline: 25770360
Fax: 22059262

Customer Assistance: