

MEDICAL EQUIPMENT REPAIR REQUEST

No. _____

Date: 05/01/2025

Requested by Shealin / 4106

Department OBG-1 Location OBG-1 opd

SR. No.	Machine	Model	Complaint
19101628	Laser machine (Fotona)	M-021-4AF/3	No laser shot coming out

Check within

 Week

 ASAP

Engineer Comments

End Comments

Shealin

Requested by

Checked By

Approved by

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



اتكو التكنولوجيه

هفون : ٢٤٧٧٧٢٥ / ٢٤٧٤٣٠١٧ (+٩٦٥)

فاكس : ٢٤٥٦٧٧٩٢ / ٢٤٧٤٣٠١٩ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63800

Customer	Al-F Hospital Jahra	Equipment	G-Runner
		Model No.	G-22
		Serial No.	21000619
Authorized Person		Manufacturer	Pelona
Customer Complaint		Location	OB 1
		Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☒ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ _____
☐ Safety Checks	☐ Operational Error		
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		
		Time Spent	

Remarks

Taking the G-Runner unit
with all its accessories to
company for detailed checking.

No.	Spare Parts Used	Part No.	Qty.

Service Engineer

ATCO Technology Co.

Signature

Date

06/01/2015

Authorized Person

Signature

Date

Customer / User

Signature

Date

9-1-25

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهره , ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63798

Customer	Alor Hospital Jahra			Equipment	Cynae Laser		
				Model No.	SPDynamis		
				Serial No.	19101628		
Authorized Person				Manufacturer	Fobong		
Customer Complaint				Location	OB 1		
				Contract / Order No.			
Services / Check Carried out ☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check ☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐				☐ Installation			☐ Warranty
				☐ Repair			☐ On Charge
				☐ Maintenance			☐ Contract
				☐ Upgrade			☐
				Time Spent			
Remarks Returned the C-Runner scanner with taken accessories. Found problem with the interface, given standby interface for working of interface standby - SN: 24000785 Old interface - SN: 24000771							
No.	Spare Parts Used			Part No.	Qty.		
Service Engineer Signature Date				Authorized Person Signature Date		Customer / User Signature Date	
Suriya ATCO Technology Co. 08/01/2025						Hassan 9-1-25	