

Wanba



MEDICAL EQUIPMENT REPAIR REQUEST

No. # 3

Date: 07-02-2024

Requested by: Feri Jy Williamson

Department:

Location: 3rd Floor

SR. No.	Machine	Model	Complaint
5627/161006	transport Ventilator (WTAHA)	Hamilton T1	* battery low

Check within

Week

☒

ASAP

Engineer Comments

Need to replace the battery of unit, waiting for the quotation.

End Comments

Replaced 2 batteries, O2 cell, Hepa @ fan filters. unit is working properly

Feri

Requested by

Checked by

Approved by

شركة وربة للتجهيزات الطبية ذ.م.م

WARBA MEDICAL SUPPLIES CO. W.L.L.

تلفون المكتب ٢٢٩ ٦٨٦ ٠٠ - المستودع: ٢٢٢ ٤٩٦ ٥٠ - سجل تجاري: ٢٧٤٦٨ - ص.ب: ٢٦٢٦٧ الصفاة، ١٣١٢٣ الكويت
Tel. Office: 229 68 600 Store: 222 496 50 - C.R. No. : 27468 - P.O.Box 26267 Safat , 13123 Kuwait
Fax: 229 68 614 / 224 57 494 - email: warba@warbamed.com, www.warbamed.com

FIELD SERVICE REPORT

Service Report No: B 2341	Date: 09/03/24 Quarter: ☐ I ☐ II ☐ III ☐ IV
Customer: AL ORF HOSPITAL	Equipment Location: N.I.C.U
Contract No:	
Equipment Name: VENTILATOR	
Manufacturer: HAMILTON	Model: TI
Serial Number: 5627	
Type of Job ☐ Preventive Maintenance ☐ Installation ☐ Warranty Replacement ☐ Emergency Maintenance ☐ Chargeable Repair ☐ Warranty Repair	
Maintenance Carried out and Final Status of the Instrument: Replaced 2 batteries, O2 cell, HEPA & FAN Filters. Software updated from 2.2.0 to 3.0.3. Done full scale calibration. unit is working fine.	

Parts Used

No.:	Part Number:	Description	Quantity
1.		LITHIUM BATTERY	2
2.		O2 CELL	1
3.		HEPA FILTER	1
4.		FAN FILTER	1

Remarks:

Name: SARFARAZ Standley Warba Engineer Signature	Name: MARYA ROUHAB Marys End User Signature & Stamp	Authorized Signature & Stamp Biomedical Department
--	---	---