

SERVICE REPORT

NO. **2218047**

SERVICE CALL NO.

Customer's name & address Al orf Hospital		Equipment MRI	
		Serial No. 31180	
		Model No. Magnetom ESPREE	
Contact person		Manufacturer SIEMENS	
Department		Maint. Cont. Ord. No.	
Tel. No.		Inv. No. Date 29.01.2025	
INSTALLATION ☐ WARRANTY ☐ PREVENTIVE MAINTENANCE ☒ CHARGEABLE ☐ REPAIR ☐ UPGRADE ☐ UPDATE ☐			
Description of WORK DONE : Preventive maintenance was done. Test tools and QA general was performed. Syngo Backup was done. Tested and handed over system for patient scan.			
Repair Completed		REMARKS	
Yes ☒	No ☐		
Eng	Date	Time From To	HRS. Work Wait Trav.
	29/1/25		
SPARE PARTS USED:		Article No.	
Description		Stock No.	
		Qty.	
Parts to be Ordered :		Description:	
Article No.		Qty:	
Service Engineer's name & Signature SRIDHAR S		CUSTOMER's name & signature	
Date 29/1/2025		Date 29/1/2025	
APPROVED BY:		Date 29/1/2025	

Maint. Frm 1