

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 60664

Customer	Aloof Hospital			Equipment	ECG		
				Model No.			
				Serial No.	2008-20022211		
Authorized Person				Manufacturer	Trismed		
Customer Complaint	Not Printing			Location	6B		
				Contract / Order No.			
Services / Check Carried out				☐ Installation		☐ Warranty	
				☐ Repair		☐ On Charge	
☐ Inspection / Cleaning				☐ Optics Checks		☐ Maintenance	
☐ Calibration / Performance				☐ Pneumatic Checks		☐ Contract	
☐ Interface / Communication				☐ Hydraulic Checks		☐ Upgrade	
☐ Safety Checks				☐ Operational Error		☐	
☐ Electrical / Electronics				☐ Mechanical Checks			
☐ Accessories Check				☐			
				Time Spent			
Remarks							
inspected unit, found unit is working normal. Battery weak - please keep unit for charging.							
No.	Spare Parts Used			Part No.	Qty.		
Service Engineer				Authorized Person		Customer / User	
Signature				Signature		Signature	
Date				Date		Date	