

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجبراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63850

Customer	Alex Hospital Jahra	Equipment	X-Ray Machine
		Model No.	
		Serial No.	
Authorized Person		Manufacturer	Primax
Customer Complaint	App Box Error	Location	X-Ray Dept
		Contract / Order No.	
Services / Check Carried out	☐ Installation ☐ Warranty ☒ Repair ☒ On Charge ☐ Maintenance ☐ Contract ☐ Upgrade ☐ Time Spent		
Remarks	Service call attended. Checked the unit and the app box needed reset, this is due to power failure. Reset the App Box. Checked and found that the unit is working in good condition.		
No.	Spare Parts Used	Part No.	Qty.
Service Engineer	Authorized Person	Customer / User	
Sunith Jibre			
ATCO Technology Co.			
Signature	Signature	Signature	
Date	Date	Date	
8/3/25		12/03/2025	