

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63512

Customer	Al Fakh Hospital Jahra			Equipment	Endoscopy Washer & Dismantler		
				Model No.	Serie 3		
				Serial No.	25186		
Authorized Person				Manufacturer	Soluscope		
Customer Complaint				Location	Endoscopy 2nd floor		
				Contract / Order No.			
Services / Check Carried out ☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check ☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐				☐ Installation		☐ Warranty	
				☒ Repair		☒ On Charge	
				☐ Maintenance		☐ Contract	
				☐ Upgrade		☐	
				Time Spent			
Remarks Replaced the water filter sets and air filters with new Dose cycle 6 and completed. checked and found that the unit is working & in good condition							
No.	Spare Parts Used			Part No.	Qty.		
Service Engineer		Authorized Person			Customer / User		
Signature Date		Signature Date			Signature Date		
Suryth ATCO Technology Co. 01/12/2024		22020120379					