

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجهر، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63669

Customer	Alof Hospital			Equipment	ENT work station		
				Model No.	Futurer 111		
				Serial No.	111632815		
Authorized Person				Manufacturer	Entermal		
Customer Complaint				Location	Ent 5th floor		
				Contract / Order No.			
Services / Check Carried out ☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check ☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐				☐ Installation		☐ Warranty	
				☒ Repair		☐ On Charge	
				☐ Maintenance		☐ Contract	
				☐ Upgrade		☐	
				Time Spent			
Remarks Service call attended. we inspect the unit and found that the enthermo is tripped off. we repaired and reset the enthermo connected the irrigation hand piece and tested. Now there is no leakage. we kept the unit under observation.							
No.	Spare Parts Used			Part No.	Qty.		
	we found the water filter is dirty. need to replace						
Service Engineer MIDAN ATCO Technology Co. Signature Date 19/05/2025		Authorized Person 22023120704 Signature Date		Customer / User Signature Date			