

ATCO



1145

MEDICAL EQUIPMENT REPAIR REQUEST

No. _____

Date: 20-08-24Requested by ERUMA W. CABILESDepartment ENT Location ~~ENT~~ FLR (DR. RAED clinic)

SR. No.	Machine	Model	Complaint
<u>103732017</u>	<u>ENTERMED - WORK STATION</u>	<u>Futurcent III</u>	<u>Cold water temperature</u>

Check within

Week

☒

ASAP

Engineer Comments

End Comments

Requested by

Checked By

Approved by

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكوتكنولوجيه

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب. ٢٣٥٥ الجهره, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 61150

Customer Alor Hospital	Equipment ENG Work Station
	Model No. Futurent II
	Serial No. 103732017
Authorized Person	Manufacturer Entermec
Customer Complaint	Location 5th Floor Dr. Raed
water temperature low	Contract / Order No.

Services / Check Carried out

- | | |
|---|---|
| ☐ Inspection / Cleaning | ☐ Optics Checks |
| ☐ Calibration / Performance | ☐ Pneumatic Checks |
| ☐ Interface / Communication | ☐ Hydraulic Checks |
| ☐ Safety Checks | ☐ Operational Error |
| ☐ Electrical / Electronics | ☐ Mechanical Checks |
| ☐ Accessories Check | ☐ _____ |

- | | |
|---|---------------------------------|
| ☐ Installation | ☐ Warranty |
| ☒ Repair | ☐ On Charge |
| ☐ Maintenance | ☐ Contract |
| ☐ Upgrade | ☐ _____ |

Time Spent

Remarks **Service call attended.**

We inspect the unit and found that the water is not heating. It is due to the water pressure low.

We bypass the pressure switch connection and the unit start working properly.

checked and confirmed that the unit is working properly.

No.	Spare Parts Used	Part No.	Qty.

Service Engineer

Nixon
ATCO Technology Co.

Signature

Date **22/09/2024**

Authorized Person

Signature

Date

Customer / User

Signature

Date

د. كريشنا شاتيا
طبيب مسجل أنف وأذن وحنجرة
مستشفى السيف
22023120704

ATCO Technology Co.