

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



log no - 1746

شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجاهراء ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 60302

Customer	Alora Hospital	Equipment	Blood Bank Refrigerator
		Model No.	DS FS 30E
		Serial No.	FF0920DS10221
Authorized Person		Manufacturer	Desmon
Customer Complaint		Location	Laboratory
High temperature		Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☒ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ _____
☐ Safety Checks	☐ Operational Error	Time Spent	
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		

Remarks

Service call attended
we inspect the unit and found that the
compressor gas is low. we refill the gas and
checked the unit.
currently the unit temperature is at the set point
we kept unit under observation

No.	Spare Parts Used	Part No.	Qty.

Service Engineer Nixon ATCO Technology Co. Signature Date 03/05/2024	Authorized Person Signature Date	Customer / User Signature Date
--	--	--------------------------------------