

INCIDENT REPORT FORM

LOG #:

Forward Completed Form to Department Manager. One copy to Quality Management Department.

PART A. Status: ☐ Inpatient ☐ Outpatient ☐ Visitor/Family ☒ Employee ☐ Other (specify) _____

Date of Occurrence: 08-04-25 Time of Occurrence: 5 PM Occurrence Location: 5TH FLOOR ALORF HOSPITAL

If event involved employee - Staff Name: ROYENA ERIKA CABRILES Age: 27yrs Sex: F Emp#: 2844

If event involved patient - Patient Name: _____ Age: _____ Sex: _____ File#: _____

If event involved Visitor - Visitor Name: _____ Phone No.: _____

Address: _____

PART B. Occurrence Type (check boxes that apply)

INTRAVENOUS RELATED	PROCEDURE VARIANCE	HAZARD/SAFETY
☐ Central Line ☐ Contrast Media Reaction ☐ Extravasate ☐ Occluded ☐ Precipitate ☐ Site Red/Swollen ☐ Tubing Leak ☐ Expired IV Fluid ☐ Additive Related ☐ Transfusion Reaction ☐ Other _____	☐ Delays ☐ Missing Files, Notes ☐ Performed on Wrong Site ☐ Consent ☐ Patient Identification ☐ Wrong Patient ☐ Wrong Procedure ☐ Lost/Spoiled Specimen ☐ No Response ☐ Wrong File ☐ Procedure Cancellation ☐ Notes in Wrong File ☐ Inadequate/Inappropriate Entry ☐ No label/Addressograph ☐ Other _____	☐ Burn ☐ Chemical Spill ☐ Electrical Shock ☐ Fire ☐ Gas ☐ Hazardous Waste ☐ Radiation Exposure ☐ Splash ☐ Water Spill ☐ Needle prick injury ☐ Other _____
BEHAVIOR	SECURITY VARIANCE	EQUIPMENT/SUPPLIES
☐ Improper disposal of sharp ☐ AMA/Absconded ☐ Dissatisfied with care ☐ Patient Behavior ☐ Family/Visitor Behavior ☐ Staff Behavior ☐ Time In/Time Out Variation ☐ Non-Compliance ☐ Patient Complaint ☐ Other _____	☐ Damage/Loss of Property ☐ Kidnapping ☐ Missing/Misplaced Drug Keys ☐ Prescription Alteration ☐ Security Problem ☐ Suicide attempt ☐ Other _____	☐ Electrical Problem ☒ Not Working ☐ Improper Use ☒ Mechanical Problem ☐ Operator Not Qualified ☐ Wrong Equipment ☐ Missing Equipment ☐ Not Available ☐ Out of Stock ☐ Other <u>ENDOSCOPE & EMT WORK STATION</u>

ADVERSE DRUG REACTION (ADR) / MEDICATION OCCURENCE

☐ Medication missing	☐ Wrong Patient	Describe Therapy:
☐ Adverse Side effects	☐ Others:	
☐ Medication given but not charted		
☐ Medication charted but not given	*please see Intravenous Related area	
☐ Duplication/extra dose given	Describe reaction (chronological sequence):	
☐ Time variance (wrong time)		Response to Treatment/Outcome:
☐ Wrong route		
☐ Wrong dose		

Reported by Name: _____ ID No. _____ Dept. _____ Position _____

IMMEDIATE ACTION Physician Notified ☐ Yes ☐ No. Did MD see patient ☐ Yes ☐ No ☐ X-ray ☐ Suture ☐ Blood Work ☐ Medication

Assessment/
Diagnosis

Physician Name:

Dept.:

ID No.:

INCIDENT REPORT FORM

"An Incident" is any happening that is not consistent with the routine operation of the hospital or the routine care of a particular patient which adversely affects or threatens the health or life of a patient, visitor or employee. It may be an occurrence or incident which might result in an accident.

PART C. Describe the occurrence (state the facts). Use when necessary.

FACTUAL DESCRIPTION

① ENDOSCOPE MACHINE MONITOR IS NOT WORKING - NO DISPLAY

② ENT WORK STATION - WATER LEAK COMING FROM WASH TIP HOSE WHEN USED

PART D. Severity Codes (Please check one).

- ☐ Minor. No injury or minor treatment, i.e. abrasion, No increased level of care.
- ☐ Moderate. Increased length of Stay. Increased level of care.
- ☐ Major. Temporary/permanent lessening of bodily functions (sensory, motor, physiologic, or intellectual). Not related to the natural Outcome of the patient's illness, underlying condition, or disfigurement. Surgical intervention required.
- ☐ Sentinel/Catastrophic. Death/Major permanent loss of function (sensory, motor, physiologic, or intellectual), hemolytic transfusion reaction, surgery on wrong patient or body part.
- ☐ Near Miss. For potentially severe Outcome.

PART E. ACTIONS.

PREVENTIVE ACTION

ACTION TAKEN BY INVOLVED DEPARTMENT		REFERRED TO QUALITY MANAGEMENT DEPT.	
is referred to Biomed DEPT. they will arrange to be seen by the engineer.		- Referred to Biomedical team. "Root Cause Analysis" to put recommendation & prevent its recurrence.	
Dept. Head:	CHRISTINA CUEVA	Date:	8/4/2024
Quality Mgmt. Dept.	Charlotte T.B. / RN	Date:	
ID No.	1022	Department:	NURSING
		Comments:	

Department Managers: Forward Completed forms to Quality Management Department



DEPT: MSD 24/05580

PROBLEM, DIAGNOSIS AND ACTION

- * REPLACED DISPLAY CABLE FROM AIDA TO TOUCH SCREEN.
- * SIGNAL ESTABLISHED.

REMARKS	
JOB COMPLETED ☒ Yes ☐ No	AWAITING PARTS ☐ Yes ☐ No

ATC ENGINEER SIGNATURE 	USER SIGNATURE 	AUTHORIZED SIGNATURE & STAMP
---	---	---