

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهره , ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 61984

Customer	Alor Hospital	Equipment	Auto clare Dental
		Model No.	sterile clare
		Serial No.	18x 25 721
Authorized Person		Manufacturer	COMINOX
Customer Complaint		Location	Dental
		Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☒ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☒ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ _____
☐ Safety Checks	☐ Operational Error		
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		
		Time Spent	

Remarks	Service call attended. checked the machine, and found there is one fuse is broken and should be replace with new ones
---------	---

No.	Spare Parts Used	Part No.	Qty.

Service Engineer Marsa ATCO Technology Co. Signature Date 21-08-2024	Authorized Person Signature Date	Customer / User Signature Date د. عبد الرحمن السعيد عبد ربه طبيب أسنان مسجل أول الاستشارة السنية مستشفى السعدي 22019100056
--	--	--