



شركة الصرح الدولية للتجارة العامة
Al-Sareh International General Trading Co.
Medical and Scientific - Equipment Supplies - Service
P.O. Box 33133 Roudha 73452 Kuwait - Tel.: (965) 22447979 - Fax : (965) 22446969

تقرير صيانة | Service Report

No.: 8909

Date: 28/11/2024

Organization : AL ORF HOSPITAL

Site/Dept. : _____

Person in Charge : _____

Designation : _____

Tel.: _____

Type of Service :

- ☐ Warranty ☐ Preventive Maintenance
☐ Installation ☐ Demonstration
☐ Chargeable ☒ On Call
☐ Delivery

Equipment Details

Contract No.: _____

Model No.: SLE6000

Description: HIGH FREQUENCY VENTILATOR Serial No.: 6020461067

Location: NICU

Manufacturer: SLE

Type of Malfunction: OXYGEN
CELL DEFECTIVE ,

Service Carried Out: DELIVERED AND
REPLACED NEW OXYGEN
CELL. CALIBRATED AND TESTED.
LEFT THE MACHINE IN
GOOD WORKING CONDITION.

Spare Parts ☒

Accessories ☒

Consumables ☒

No.	Item	Qty	No.	Item	Qty
1.	OXYGEN CELL PIN: N2191/10	1			

Service Completed

☒ Yes ☐ No

Time Spend: Hrs: _____ Min: _____

Service Engineer

MUJEER
M.T.

Engineering Department

Customer's Signature

Fateh Jay Williams

