

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجھراء، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 60124

Customer	Aloft Hospital	Equipment	Blood Bank Refrigerator
		Model No.	DS-F30E
		Serial No.	FF0920DSK0221
Authorized Person		Manufacturer	Desman Scientific
Customer Complaint		Location	Laboratory.
		Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
		☒ Repair	☐ On Charge
		☐ Maintenance	☐ Contract
		☐ Upgrade	☐ _____
		Time Spent	
Remarks			

Services / Check Carried out

- | | |
|---|---|
| ☐ Inspection / Cleaning | ☐ Optics Checks |
| ☐ Calibration / Performance | ☐ Pneumatic Checks |
| ☐ Interface / Communication | ☐ Hydraulic Checks |
| ☐ Safety Checks | ☐ Operational Error |
| ☐ Electrical / Electronics | ☐ Mechanical Checks |
| ☐ Accessories Check | ☐ _____ |

Remarks Service call attended.

We inspect the unit and found that the refrigerator temperature is not reaching the set point.

Also there is one error alarm AL03 also come.

For further error we need to check with the manufacturer regarding the issue.

Currently the unit temperature is not in limit.

No.	Spare Parts Used	Part No.	Qty.

Service Engineer Nicom ATCO Technology Co. Signature Date 30/04/2024	Authorized Person Signature Date	Customer / User Signature Date 2-5-24
--	--	---