

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



log no - 2592

شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 61083

Customer	Alert Hospital Jahra			Equipment	Steam Sterilizer		
				Model No.	Sechebmat		
				Serial No.	220824		
Authorized Person				Manufacturer	MMM		
Customer Complaint	Bowie Dwb Failure			Location	CSSD		
				Contract / Order No.			
Services / Check Carried out				☐ Installation	☐ Warranty		
				☒ Repair	☐ On Charge		
☐ Inspection / Cleaning				☐ Maintenance	☐ Contract		
☐ Calibration / Performance				☐ Upgrade	☐ _____		
☐ Interface / Communication				Time Spent			
☐ Safety Checks							
☐ Electrical / Electronics							
☐ Accessories Check							
Remarks							
Service call attended. Checked the unit and found the both chillers cold water is not present. Checked on the top and found the both chillers are not working and hot water in the tank. Please solve the problem							
No.	Spare Parts Used		Part No.	Qty.			
	for the chiller for the proper working of the auto clave.						
	under warranty						
Service Engineer		Authorized Person		Customer / User			
Signature		Signature		Signature			
Date		Date		Date			
25/08/2024		26/8/24					