

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجهرام، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 58651

Customer	Allof Hognik	Equipment	Patent Monitor
		Model No.	Vitapia 7000
		Serial No.	05BM000271
Authorized Person		Manufacturer	Drismed
Customer Complaint		Location	LDR
	SpO2 not working	Contract / Order No.	

Services / Check Carried out	☐ Installation	☐ Warranty
	☐ Repair	☐ On Charge
☐ Inspection / Cleaning	☐ Optics Checks	☐ Maintenance
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade
☐ Safety Checks	☐ Operational Error	☐
☐ Electrical / Electronics	☐ Mechanical Checks	
☐ Accessories Check	☐	
	Time Spent	

Remarks	Inspected unit found new SpO2 sensors not working with this unit. Please check sensors are for this unit or not.
	Tested unit with other SpO2 sensor and found working ok

No.	Spare Parts Used	Part No.	Qty.

Service Engineer	Authorized Person	Customer / User
ATCO Technology Co.		
Signature	Signature	Signature
Date	Date	Date
18/2/24		19-2-24