

P.O.Box : 24240 Safat

Code : 13103 | Kuwait

C.R Number : 353520

Capital 500,000 K.D



شركة السلطان المتحدة الطبية
AlSultan United Medical Co.

ص.ب : 24240 الصفاة

الرمز: 13103 | الكويت

السجل التجاري : 353520

رأس المال 500,000 د.ك

SERVICE REPORT

NO. 2218677

SERVICE CALL NO.

Customer's name & address <u>Mansoor Alorf Hospital</u>				Equipment <u>CT</u>			
				Serial No. <u>59412</u>			
				Model No. <u>Somatom Perspective</u>			
Contact person _____				Manufacturer <u>Siemens</u>			
Department _____				Maint. Cont. _____ Ord. No. _____			
Tel. No. _____				Inv. No. _____ Date <u>10.03.2025</u>			

INSTALLATION ☐ WARRANTY ☐ PREVENTIVE MAINTENANCE ☐ CHARGEABLE ☒ REPAIR ☒ UPGRADE ☐ UPDATE ☐

Description of WORK DONE : Preventive maintenance done.
2 piece P68 gantry fans replaced with new ones.
Gantry filter (10856949), detector filter (10186056) replaced.
Detector fan (07071921) replaced. Signal brush (10415047) replaced.
Gantry and table greasing done.
Quality constancy and calibrations done.
System is tested and given to the customer in working condition.

Repair Completed Yes ☒ No ☐							REMARKS	
Eng	Date	Time		HRS.			Over	
		From	To	Work	Wait	Trav.	Hrs.	
ABD 03 0A	10.03.2025							

SPARE PARTS USED:		Article No.	Stock No.	Qty.
Description				
P68 Gantry Fan		10891306		2
P68 Air filter sponge set		10856949		1
Filter mat		10186056		1
FAN 24V (Detector fan)		07071921		1
Signal-brush-FG		10415047		1
				7

Parts to be Ordered :	Article No.	Description:	Qty:

Service Engineer's name & Signature <u>Abdullah Safak</u>		CUSTOMER's name & signature <u>DUA ABDELLAH</u>		APPROVED BY: <u>DUA ABDELLAH</u>	
Date <u>10.03.2025</u>		Date		Date	

Maint. Frm 1