

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهره , ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 62617

Customer	Alor Hospital	Equipment	Defibrillator.
		Model No.	Cardio Aid 360-B
		Serial No.	212060645
Authorized Person		Manufacturer	Innomed.
Customer Complaint		Location	4th floor (CARDIO)
Battery faulty		Contract / Order No.	
Services / Check Carried out		☐ Installation ☐ Warranty	
☐ Inspection / Cleaning		☒ Repair	
☐ Calibration / Performance		☐ On Charge	
☐ Interface / Communication		☐ Maintenance	
☐ Safety Checks		☐ Contract	
☐ Electrical / Electronics		☐ Upgrade	
☐ Accessories Check		☐	
		Time Spent	
Remarks			
we replaced the defibrillator battery with new one. Tested the daily test and functions and found that the unit is working properly.			
checked and confirmed that the unit is working properly.			
No.	Spare Parts Used	Part No.	Qty.
1.	CA 360B Battery Pack		1
	SN / MS 3870056		
Service Engineer		Authorized Person	
Nixon		د. محمد عبد الله	
ATCO Technology Co.		استشاري امراض القلب	
Signature		Signature	
Date		Date	
16/07/2024		12/07/2024	
		Customer / User	
		Signature	
		Date	