

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجھراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 61169

Customer	ALORF HOSPITAL	Equipment	Microscope
	Jahra	Model No.	CX41 RF
		Serial No.	4C08854
Authorized Person		Manufacturer	Olympus
Customer Complaint		Location	IVF
	Lamp not working	Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☐ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ EM
☐ Safety Checks	☐ Operational Error		
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐		
		Time Spent	

Remarks

Re connected and re fixed the lamp.
Insulated the lamp cables. Cleaned the lamp connectors
All connections are tested ok
Microscope working good

No.	Spare Parts Used	Part No.	Qty.

Service Engineer

Jince Cherian
ATCO Technology Co.

Signature

Date

15/05/2025

Authorized Person

Signature

Date

Customer / User

MIRA / 15-5-25

Signature

Date