



log no-1781

شركة ماجلان للتجارة العامة والمقاولات ذ.م.م.
أجهزة طبية وعلمية وأدوات مخبرية

قسم الصيانة
SERVICE REPORT

8139

Customer : Albst Hospital Date: 22/10/24

Department: Radiology

Type of equipment : DNRs Bone densitometer

Model No.: Stratus Serial No.: 211 015D256

Service Call Type :

installation

Repair

Maintenance

Emergency

☐☒☐☐

Warranty

Contract

Chargeable

Others

☐☐☐☐

ENGINEER'S REPORT : Fault : Samarium of Detector has been

Action : replaced.

Software has been Upgraded.

Unit Calibration is performed.

Remarks : Unit found to be working

Warranty : good.

Repair Time : _____

Parts Used : _____

Invoice No.: _____ Date : _____

K.D.

Fils

ENGINEER'S NAME AND SIGNATURE

CUSTOMER'S NAME AND SIGNATURE