

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



log no - 2591 شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 60983

Customer	Alor Hospital		
Equipment	Autoclave		
Model No.	Selectomab PL		
Serial No.	220823		
Authorized Person	Manufacturer		
Customer Complaint	Location		
spots on outer covering of instruments			
Contract / Order No.			
☐ Installation ☐ Warranty			
☒ Repair ☐ On Charge			
☐ Maintenance ☐ Contract			
☐ Upgrade ☐			
Time Spent			
Services / Check Carried out			
☐ Inspection / Cleaning ☐ Optics Checks			
☐ Calibration / Performance ☐ Pneumatic Checks			
☐ Interface / Communication ☐ Hydraulic Checks			
☐ Safety Checks ☐ Operational Error			
☐ Electrical / Electronics ☐ Mechanical Checks			
☐ Accessories Check ☐			
Remarks			
Service call attended. we inspect the unit and suspecting the issue due to the water quality for Steam generator. we cleaned the Chamber using Sand Paper to remove the Particles Settled. The water Softener need Service and its chargeable. The autoclave unit is working, only the Steam generator water quality is the issue.			
No.	Spare Parts Used	Part No.	Qty.
under warranty			
Service Engineer			
Authorized Person			
Customer / User			
ATCO Technology Co.			
Signature			
Date			
15/07/2024			