

Customer Aloof Hospital		Equipment Autoclave	
		Model No. selectomat PL	
		Serial No. 220823	
Authorized Person		Manufacturer mmm	
Customer Complaint		Location CSSD	
Stain on equipment.		Contract / Order No.	
Services / Check Carried out ☐ Inspection / Cleaning ☐ Optics Checks ☐ Calibration / Performance ☐ Pneumatic Checks ☐ Interface / Communication ☐ Hydraulic Checks ☐ Safety Checks ☐ Operational Error ☐ Electrical / Electronics ☐ Mechanical Checks ☐ Accessories Check ☐ _____		☐ Installation ☐ Warranty	
		☒ Repair ☐ On Charge	
		☐ Maintenance ☐ Contract	
		☐ Upgrade ☐ _____	
		Time Spent	
Remarks Service call attended. we inspect the unit and cleaned the steam generator using pressure pump. we found that the impurities settle down in the steam generator now the unit is working properly checked and confirmed that the unit is working properly			
No.	Spare Parts Used	Part No.	Qty.
Service Engineer Nidoni/Smith/Manu ATCO Technology Co.		Authorized Person 18/8/24	
Signature 		Signature	
Date 15/08/2022		Date	
Customer / User		Signature	
		Date	