

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجهرام, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63007

Customer	Mory Hospital Jahra	Equipment	OT - Table
		Model No.	V 750
		Serial No.	NV111020091
Authorized Person		Manufacturer	NUVO
Customer Complaint		Location	OT - Room 5
		Contract / Order No.	

Services / Check Carried out

- | | |
|--|--|
| ☐ Inspection / Cleaning | ☐ Optics Checks |
| ☐ Calibration / Performance | ☐ Pneumatic Checks |
| ☐ Interface / Communication | ☐ Hydraulic Checks |
| ☐ Safety Checks | ☐ Operational Error |
| ☐ Electrical / Electronics | ☐ Mechanical Checks |
| ☐ Accessories Check | ☐ _____ |

- | | |
|---------------------------------------|------------------------------------|
| ☐ Installation | ☐ Warranty |
| ☐ Repair | ☐ On Charge |
| ☐ Maintenance | ☐ Contract |
| ☐ Upgrade | ☐ _____ |

Time Spent

Remarks

Oil leaking
Found leak from the base panel.
clean the base panel and tubing.

No.	Spare Parts Used	Part No.	Qty.
	oil		
	still LEAKED.		

Service Engineer

Imine Cherian
ATCO Technology Co.

Signature

Date

23/04/25

Authorized Person

Signature

Date

20/5/25

Customer / User

Signature

Date