


Customer Aloof Hospital		Equipment Autoclave	
		Model No. Selectom at PL	
		Serial No. 220823 & 220824	
Authorized Person		Manufacturer MMM	
Customer Complaint Water Temperature Error		Location CSSD	
		Contract / Order No.	
Services / Check Carried out ☐ Inspection / Cleaning ☐ Optics Checks ☐ Calibration / Performance ☐ Pneumatic Checks ☐ Interface / Communication ☐ Hydraulic Checks ☐ Safety Checks ☐ Operational Error ☐ Electrical / Electronics ☐ Mechanical Checks ☐ Accessories Check ☐		☐ Installation ☐ Warranty	
		☒ Repair ☐ On Charge	
		☐ Maintenance ☐ Contract	
		☐ Upgrade ☐	
		Time Spent	
Remarks Service call attended, we inspect the unit and found that there is no cool water supply for both the unit. So the error happen. We inform the maintenance team and they solve the cool water supply. Now the unit is working. Checked and confirmed that the unit is working properly.			
No.	Spare Parts Used	Part No.	Qty.
Service Engineer Niranjan ATCO Technology Co. Signature  Date 23/04/2024		Authorized Person  Signature Date	
		Customer / User Signature Date	