



MEDICAL EQUIPMENT REPAIR REQUEST

No. _____

Date: 05/05/25

Requested by Amr Abdel Antony

Department PT Location _____

SR. No.	Machine	Model	Complaint
<u>NA</u>	<u>Bed</u>	<u>Medisa</u> <u>(Atco tech)</u>	<u>Pedal is not working</u>

Check within Week

☒ ASAP

Engineer Comments

End Comments

Amr

Requested by

Checked By

Approved by

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهره , ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63015

Customer	ALORF Hospital	Equipment	Observation bed
	Jahra	Model No.	
		Serial No.	
Authorized Person		Manufacturer	Meelisa
Customer Complaint	Foot switch not working	Location	PT
		Contract / Order No.	-

Services / Check Carried out	☐ Installation	☐ Warranty
	☐ Repair	☐ On Charge
☐ Inspection / Cleaning	☐ Maintenance	☐ Contract
☐ Calibration / Performance	☐ Upgrade	☐ _____
☐ Interface / Communication	Time Spent	
☐ Safety Checks		
☐ Electrical / Electronics		
☐ Accessories Check		
☐ Optics Checks		
☐ Pneumatic Checks		
☐ Hydraulic Checks		
☐ Operational Error		
☐ Mechanical Checks		
☐ _____		

Remarks
checked found foot switch not working
Take foot switch taken to workshop for repair.

No.	Spare Parts Used	Part No.	Qty.

Service Engineer Sirine Cherian ATCO Technology Co. Signature Date 08/05/25	Authorized Person د. نادية أحمد محمد مسجل الطب الطبيعي والتأهيل مستشفى العرف 133161983 Signature Date	Customer / User Anumol Antony Signature Date 08/05/25
---	---	--