

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



log no - 2591

شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء , ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 60320

Customer	Aloof Hospital	Equipment	Autoclave
		Model No.	Selectomab PL
		Serial No.	220823
Authorized Person		Manufacturer	MMM
Customer Complaint		Location	CSSD
		Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☒ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ _____
☐ Safety Checks	☐ Operational Error		
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		
		Time Spent	

Remarks Service call attended, we inspect the unit and found that the cold water from the chiller is not cool. The high temperature of cold water effect the vacuum. Need to upgrade the chiller for the proper working of autoclave machine.

No.	Spare Parts Used	Part No.	Qty.

Service Engineer Nixon ATCO Technology Co. Signature Date 22/06/2024.	Authorized Person Signature Date 23/6/24	Customer / User Signature Date
---	--	--------------------------------------